



Recent Data Anomalies at USGS Site 01651000 February 2009 NW Branch Anacostia River near Hyattsville, MD

USGS Real-Time Data at this site

State standards: Temperature $\leq 32.2^{\circ}\text{C}$ Turbidity ≤ 150 NTU $6.5 \leq \text{pH} \leq 8.5$ Dissolved oxygen > 5.0 mg/l

From June 30, 2007, until February 15, 2008, water-quality data collection at this streamgage was discontinued due to funding reductions from local agencies.

Parameter	Date	Value
Turbidity	2009-02-01 12:30	170
Turbidity	2009-02-01 12:45	220
Turbidity	2009-02-01 13:00	230
Turbidity	2009-02-01 13:15	250
Turbidity	2009-02-01 13:30	260
Turbidity	2009-02-01 13:45	300
Turbidity	2009-02-01 14:00	280
Turbidity	2009-02-01 14:15	290
Turbidity	2009-02-01 14:30	310
Turbidity	2009-02-01 14:45	330
Turbidity	2009-02-01 15:00	320
Turbidity	2009-02-01 15:15	300
Turbidity	2009-02-01 15:30	310
Turbidity	2009-02-01 15:45	310
Turbidity	2009-02-01 16:00	300
Turbidity	2009-02-01 16:15	310
Turbidity	2009-02-01 16:30	290
Turbidity	2009-02-01 16:45	290
Turbidity	2009-02-01 17:00	280
Turbidity	2009-02-01 17:15	280
Turbidity	2009-02-01 17:30	270
Turbidity	2009-02-01 17:45	260
Turbidity	2009-02-01 18:00	250
Turbidity	2009-02-01 18:15	230
Turbidity	2009-02-01 18:30	220
Turbidity	2009-02-01 18:45	200

Parameter	Date	Value
Turbidity	2009-02-01 19:00	200
Turbidity	2009-02-01 19:15	170
Turbidity	2009-02-01 19:30	160
pH	2009-02-03 14:30	1730
pH	2009-02-03 14:45	1730
pH	2009-02-03 15:00	1730
pH	2009-02-10 11:00	779
pH	2009-02-17 08:45	891