



Recent Data Anomalies at USGS Site 01651000 February 2010 NW Branch Anacostia River near Hyattsville, MD

USGS Real-Time Data at this site

State standards: Temperature $\leq 32.2^{\circ}\text{C}$ Turbidity ≤ 150 NTU $6.5 \leq \text{pH} \leq 8.5$ Dissolved oxygen > 5.0 mg/l

From June 30, 2007, until February 15, 2008, water-quality data collection at this streamgage was discontinued due to funding reductions from local agencies.

Parameter	Date	Value
pH	2010-02-11 15:45	5930
pH	2010-02-11 16:00	6020
pH	2010-02-11 16:15	6140
pH	2010-02-11 16:30	6270
pH	2010-02-23 09:30	1140
pH	2010-02-23 09:45	1140
pH	2010-02-23 10:00	1140
pH	2010-02-23 10:30	1140
pH	2010-02-23 10:45	1150