



# Recent Data Anomalies at USGS Site 01651000 January 2009 NW Branch Anacostia River near Hyattsville, MD

## USGS Real-Time Data at this site

State standards: Temperature  $\leq 32.2^{\circ}\text{C}$  Turbidity  $\leq 150$  NTU  $6.5 \leq \text{pH} \leq 8.5$  Dissolved oxygen  $> 5.0$  mg/l

From June 30, 2007, until February 15, 2008, water-quality data collection at this streamgage was discontinued due to funding reductions from local agencies.

Parameter	Date	Value
pH	2009-01-01 19:00	405
pH	2009-01-01 19:15	403
pH	2009-01-01 20:00	402
pH	2009-01-01 20:15	402
pH	2009-01-01 21:00	402
pH	2009-01-01 21:15	402
pH	2009-01-01 22:00	401
pH	2009-01-01 22:15	401
pH	2009-01-01 23:00	401
pH	2009-01-01 23:15	401
Turbidity	2009-01-07 10:30	270
Turbidity	2009-01-07 16:30	160
Turbidity	2009-01-10 21:00	410
pH	2009-01-13 12:45	402
pH	2009-01-13 13:00	403
pH	2009-01-13 13:15	404
pH	2009-01-13 13:30	404
pH	2009-01-13 13:45	403
pH	2009-01-13 14:00	403
pH	2009-01-13 14:15	403
Turbidity	2009-01-28 16:45	170