



# Recent Data Anomalies at USGS Site 01651000 July 2009

## NW Branch Anacostia River near Hyattsville, MD

### USGS Real-Time Data at this site

State standards: Temperature  $\leq 32.2^{\circ}\text{C}$  Turbidity  $\leq 150$  NTU  $6.5 \leq \text{pH} \leq 8.5$  Dissolved oxygen  $> 5.0$  mg/l

From June 30, 2007, until February 15, 2008, water-quality data collection at this streamgage was discontinued due to funding reductions from local agencies.

Parameter	Date	Value
pH	2009-07-10 10:15	21.0
pH	2009-07-10 11:45	441
pH	2009-07-15 10:45	448
pH	2009-07-15 14:00	447
pH	2009-07-15 14:15	447
pH	2009-07-15 14:30	447
pH	2009-07-15 14:45	447
pH	2009-07-15 15:00	447
pH	2009-07-15 15:15	447
pH	2009-07-15 15:30	447
pH	2009-07-15 15:45	447
pH	2009-07-15 16:00	446
pH	2009-07-15 16:15	446
pH	2009-07-15 16:30	446
pH	2009-07-15 16:45	446
pH	2009-07-15 17:00	447
pH	2009-07-15 17:15	447
pH	2009-07-15 17:30	447
pH	2009-07-15 17:45	447
pH	2009-07-15 18:00	448
pH	2009-07-15 18:15	448
pH	2009-07-15 18:30	448
pH	2009-07-15 18:45	449
pH	2009-07-15 19:00	449
pH	2009-07-15 19:15	449
pH	2009-07-15 19:30	449

<b>Parameter</b>	<b>Date</b>	<b>Value</b>
pH	2009-07-15 19:45	449
pH	2009-07-15 20:00	449
pH	2009-07-15 20:15	449
pH	2009-07-15 20:30	449
pH	2009-07-15 20:45	449
pH	2009-07-15 21:00	449
pH	2009-07-15 21:15	449
pH	2009-07-15 21:30	449
pH	2009-07-15 21:45	449
pH	2009-07-15 22:00	449
pH	2009-07-15 22:15	449
pH	2009-07-15 22:30	449
pH	2009-07-15 22:45	449
pH	2009-07-15 23:00	449
pH	2009-07-15 23:15	449
pH	2009-07-15 23:30	448
pH	2009-07-15 23:45	448
pH	2009-07-16 00:00	448
pH	2009-07-16 00:15	448
pH	2009-07-16 00:30	448
pH	2009-07-16 01:15	23.4
pH	2009-07-16 01:30	23.3
pH	2009-07-16 01:45	447
pH	2009-07-16 02:00	447
pH	2009-07-16 02:15	447
pH	2009-07-16 02:30	447
pH	2009-07-16 02:45	447
pH	2009-07-16 03:00	447
pH	2009-07-16 03:15	447
pH	2009-07-16 03:30	447
pH	2009-07-16 03:45	447
pH	2009-07-16 04:00	447
pH	2009-07-16 04:15	448
pH	2009-07-16 04:30	448
pH	2009-07-16 04:45	448
pH	2009-07-16 05:00	448
pH	2009-07-16 11:15	448
pH	2009-07-16 11:30	448
Turbidity	2009-07-23 23:15	250

<b>Parameter</b>	<b>Date</b>	<b>Value</b>
Turbidity	2009-07-23 23:30	210
Turbidity	2009-07-23 23:45	210
Turbidity	2009-07-24 00:00	190
Turbidity	2009-07-24 00:15	160
Dissolved Oxygen	2009-07-28 00:00	4.9
Dissolved Oxygen	2009-07-28 01:00	5.0
Dissolved Oxygen	2009-07-28 01:15	5.0
Dissolved Oxygen	2009-07-28 01:30	4.9
Dissolved Oxygen	2009-07-28 01:45	5.0
Dissolved Oxygen	2009-07-28 02:00	5.0
Dissolved Oxygen	2009-07-28 02:15	5.0
Dissolved Oxygen	2009-07-28 02:45	5.0
Dissolved Oxygen	2009-07-28 03:00	5.0
pH	2009-07-28 18:15	340
pH	2009-07-28 21:15	344
Dissolved Oxygen	2009-07-30 03:45	5.0
Dissolved Oxygen	2009-07-30 04:45	5.0
Dissolved Oxygen	2009-07-30 05:30	5.0
Dissolved Oxygen	2009-07-30 05:45	5.0
Dissolved Oxygen	2009-07-30 06:00	5.0
Dissolved Oxygen	2009-07-30 06:30	5.0
pH	2009-07-30 07:45	336
pH	2009-07-30 08:00	337
pH	2009-07-30 08:15	338
pH	2009-07-30 08:30	339
pH	2009-07-30 08:45	340
pH	2009-07-30 09:00	341
pH	2009-07-30 09:15	341
pH	2009-07-30 09:30	342