



# Recent Data Anomalies at USGS Site 01651000

## November 2009

### NW Branch Anacostia River near Hyattsville, MD

#### USGS Real-Time Data at this site

State standards: Temperature  $\leq 32.2^{\circ}\text{C}$  Turbidity  $\leq 150$  NTU  $6.5 \leq \text{pH} \leq 8.5$  Dissolved oxygen  $> 5.0$  mg/l

From June 30, 2007, until February 15, 2008, water-quality data collection at this streamgage was discontinued due to funding reductions from local agencies.

Parameter	Date	Value
pH	2009-11-02 11:30	252
pH	2009-11-02 11:45	242
pH	2009-11-02 12:00	242
pH	2009-11-02 12:15	244
Dissolved Oxygen	2009-11-02 13:30	3.6
pH	2009-11-02 13:30	244
Turbidity	2009-11-19 23:30	340
Turbidity	2009-11-19 23:45	290
Turbidity	2009-11-20 00:00	660
Turbidity	2009-11-20 00:15	250
Turbidity	2009-11-20 00:30	210
Turbidity	2009-11-20 00:45	230
Turbidity	2009-11-20 01:00	190
Turbidity	2009-11-20 01:15	180
Turbidity	2009-11-20 01:30	180
Turbidity	2009-11-20 01:45	170
Turbidity	2009-11-20 02:00	160
Turbidity	2009-11-20 02:15	160
Turbidity	2009-11-20 02:30	160
Turbidity	2009-11-20 06:00	180
Turbidity	2009-11-20 06:15	200
Turbidity	2009-11-20 06:30	210
Turbidity	2009-11-20 06:45	200
Turbidity	2009-11-20 07:00	190
Turbidity	2009-11-20 07:15	180
Turbidity	2009-11-20 07:30	170

<b>Parameter</b>	<b>Date</b>	<b>Value</b>
Turbidity	2009-11-20 07:45	160
Turbidity	2009-11-20 08:00	160
Turbidity	2009-11-20 08:15	160
Turbidity	2009-11-20 08:30	160
Turbidity	2009-11-20 08:45	180
Turbidity	2009-11-20 09:00	170
Turbidity	2009-11-20 09:15	170
Turbidity	2009-11-20 09:30	170
Turbidity	2009-11-20 09:45	160
Turbidity	2009-11-20 10:00	160
Turbidity	2009-11-20 10:15	170
Turbidity	2009-11-20 10:30	170
Turbidity	2009-11-20 10:45	180
Turbidity	2009-11-20 11:00	180
Turbidity	2009-11-20 11:15	180
Turbidity	2009-11-20 11:30	170
Turbidity	2009-11-20 11:45	170
Turbidity	2009-11-20 12:00	170